

CAMPDEN BRI RETAILER SUPPLEMENTARY AUDIT SCOPE OF COMPLIANCE

Following the audit of Advanced Laboratory Testing, Newbridge, conducted on 24 November 2020, compliance has been awarded for the following tests.

Date of issue: 24 December 2020
Scope valid until: 28 February 2022
Annual surveillance RSA visit due: November 2021

Test Title	Lab Method Ref No.	Source Method
Food Matrices		
<u>Enumeration of:</u>		
Coagulase Positive <i>Staphylococcus aureus</i>	MTM014	ISO6888-1:1999+AMD 2:2018 with confirmation by latex test (DiaMondiaL Staph Plus)
<i>Clostridium perfringens</i>	MTM018	ISO7937:2004 with confirmation by biochemical gallery (Rapid 32A)
Presumptive <i>Bacillus cereus</i>	MTM015	ISO 7932:2005
<i>Listeria</i> spp. and <i>Listeria monocytogenes</i>	MTM016 + MTM006	ISO 11209-2:2017 incorporating confirmation by biochemical gallery (Microgen Listeria ID)
<i>Campylobacter</i> spp.	MTM041	ISO10272-2:2017
<u>Detection of:</u>		
<i>Salmonella</i> spp.	MTM063 + MTM003	BioRad Rapid Salmonella short protocol AFNOR BRD 07-/11-12/05 with confirmation by ISO 6579-1:2007+Amd1:2020 (oxidase, serology and biochemical gallery (API20E)
<i>Listeria</i> spp. and <i>Listeria monocytogenes</i>	MTM045 + MTM006	Oxoid Listeria Precis AFNOR UNI 03/04-04/05 with confirmation by ISO 11290-1:2017 (Gram stain, catalase, oxidase, haemolysis and biochemical gallery - Microgen Listeria ID)
<i>Campylobacter</i> spp. (Materials with low Campylobacter numbers and low numbers of background microflora and/or stressed Campylobacter e.g. cooked and frozen materials. Materials with low numbers of Campylobacter and high levels of background microflora e.g. raw meats (incl. poultry), raw milk)	MTM007	ISO10272-1:2017 (Enrichment Options A and B)

Test Title	Lab Method Ref No.	Source Method
Environmental Swabs		
<u>Enumeration of:</u>		
Coagulase Positive <i>Staphylococcus aureus</i>	MTM014	ISO6888-1:1999+AMD 2:2018 with confirmation by latex test (DiaMondiaL Staph Plus)
<i>Listeria</i> spp. and <i>Listeria monocytogenes</i>	MTM016 + MTM006	ISO 11209-2:2017 incorporating confirmation by biochemical gallery (Microgen Listeria ID)
<u>Detection of:</u>		
<i>Salmonella</i> spp.	MTM063 + MTM003	BioRad Rapid Salmonella Short protocol AFNOR BRD 07-/11-12/05 with confirmation by ISO 6579-1:2007+Amd1:2020 (oxidase, serology and biochemical gallery (API20E)
<i>Listeria</i> spp. and <i>Listeria monocytogenes</i>	MTM045 + MTM006	Oxoid Listeria Precis AFNOR UNI 03/04-04/05 with confirmation by ISO 11290-1:2017 (Gram stain, catalase, oxidase, haemolysis and biochemical gallery - Microgen Listeria ID)